

## CE Declaration of Conformity Type Ringcaps

We hereby confirm, that the product

### Type Ringcaps

(Productnumber: see annex)

corresponds to risk class A according to Annex VIII 2.5 rule 5 c).


The conformity assessment procedure according to Article 48 (10) was carried out according to Annex I, II and III of the IVDR (EU) 2017/746.

<b>Manufacturer:</b>	Hirschmann Laborgeräte GmbH & Co. KG Hauptstr. 7-15 · 74246 Eberstadt/Germany
<b>Single Registration Number (SRN):</b>	DE-MF-000028167
<b>Basic UDI-DI:</b>	4049338DECA37IVD12E2
<b>Intended use :</b>	Sample receptable for capillary blood collection

The manufacturer is solely responsible for the issuance of the certificate of conformity. We declare that the products covered by this declaration comply with this Regulation and, where applicable, with other relevant Union legislation providing for the issuance of an EU declaration of conformity.


The declaration of conformity is valid until a change is made to the product, at the latest until 26/05/2027.

20/01/2023



Kirsten Hirschmann  
CEO // Responsible person  
for regulatory compliance

20/01/2023



Marc Bangert  
Prokurist/Deputy // responsible  
person for regulatory compliance

HIRSCHMANN®

Hirschmann Laborgeräte GmbH & Co. KG  
Hauptstraße 7-15  
74246 Eberstadt | Germany

Company stamp

Rev. 01/2023

Page 1 of 2

## Annex Type Ringcaps

---

The following product numbers are summarized under the product group Ringcaps:

No.	Description
9600105	RINGCAPS 1+2+3+4+5 µl
9600110	RINGCAPS 10µl
9600120	RINGCAPS 20 µl
9600122	RINGCAPS 200 µl
9600125	RINGCAPS 25 µl
9600140	RINGCAPS 40 µl
9600144	RINGCAPS 44,7 µl
9600150	RINGCAPS 50 µl
9600199	RINGCAPS 100 µl
9600210	RINGCAPS 5+10 µl
9600222	RINGCAPS 100+200 µl
9600240	RINGCAPS 20+40 µl
9600299	RINGCAPS 50+100 µl
9602144	RINGCAPS 44,7 µl, NA-HEP